



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/825,128	04/16/2004	Carl-Goran Hagert	HW-8645	7993
26294 7590 10/01/2009 TAROLLI, SUNDHEIM, COVELL & TUMMINO L.L.P. 1300 EAST NINTH STREET, SUITE 1700 CLEVEVLAND, OH 44114				
EXAMINER				
ARAJ, MICHAEL J				
ART UNIT		PAPER NUMBER		
3775				
MAIL DATE		DELIVERY MODE		
10/01/2009		PAPER		

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

### Office Action Summary

**Application No.**

10/825,128

**Applicant(s)**

HAGERT, CARL-GORAN

**Examiner**

MICHAEL J. ARAJ

**Art Unit**

3775

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 22 January 2009.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 11, 13, 15-29, 31 and 32 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 11, 13, 15-29, 31 and 32 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some \* c) ☐ None of:
1. ☒ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB-08)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: \_\_\_\_\_

## DETAILED ACTION

### *Claim Objections*

Claims 15-17 and 22 objected to because of the following informalities: Claims 15 and 22 are dependent upon claim 14 which has been cancelled. Since claim 17 is dependent upon claim 16 and claim 16 is dependent upon claim 15, these claims are also objected to. **Appropriate correction is required.** For examination purposes, claim 15 and 22 will be observed as being dependent upon claim 11.

### *Claim Rejections - 35 USC § 102*

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 11, 15-18, 20-29, 31 and 32 is rejected under 35 U.S.C. 102(b) as being anticipated by Murray (U.S. Patent No. 3,877,424).

Murray discloses a system comprising a polyethylene chassis (18) [Col 2. line34-37] and one or more one-piece fixation elements (8) in the form of pins adapted to be received in a bone structure. The fixation elements are received in the polyethylene chassis wherein the polyethylene chassis in effect locks the fixation elements by friction to prevent axial, rotational and angular movement of the fixation element relative to the chassis. The fixation elements are to be received in a bone structure and are screwed into the chassis and bone structure in such a way that screws move equidistantly in the

chassis and the bone structure to produce a force neutral form. There are also no axial forces transferred to the screws after fixation. The chassis is received in a rigid bracing (epoxy resin) wherein the chassis is made of two parts (Fig. 7) received displaceable in an axial direction in relation to each other in the bracing that a gap is formed between the two chassis parts. This device is placed at a distance from and not in contact with the underlying bone structure or skin (see figure 1). The chassis is fixed to both sides of the fracture whereby a bridge span is formed between bone fragments of the bone.

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 13 and 19 rejected under 35 U.S.C. 103(a) as being unpatentable over Murray (U.S. Patent No. 3,877,424).

Murray discloses the claimed invention except for the chassis being made of ultra high molecular weight polyethylene or the bracing being made of steel. It would have been obvious to one having ordinary skill in the art at the time the invention was made to have created the chassis with ultra high molecular weight polyethylene or the bracing rod of steel, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. In re Leshin, 125 USPQ 416.

### ***Response to Arguments***

Applicant's arguments with respect to claims 11, 13, 15-19, 20-29 and 31 have been considered but are moot in view of the new ground(s) of rejection. The amendment of "said fixation elements are frictionally engaged by said polyethylene chassis" appears to overcome the Weisenbach reference because the fixation element (8) makes direct contact with the nut (26) and not "direct" contact with the chassis (12).

### ***Conclusion***

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MICHAEL J. ARAJ whose telephone number is (571)272-5963. The examiner can normally be reached on M-F 8am-5pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Thomas Barrett can be reached on 571-272-4746. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

/Michael J Araj/  
Examiner, Art Unit 3775

/Thomas C. Barrett/  
Supervisory Patent Examiner, Art Unit 3775